



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/788,966	02/25/2004	Leonard Pinchuk	BSI-430US10	8808
23122	7590	06/21/2006	EXAMINER	
RATNERPRESTIA P O BOX 980 VALLEY FORGE, PA 19482-0980			PELLEGRINO, BRIAN E	
			ART UNIT	PAPER NUMBER
			3738	
DATE MAILED: 06/21/2006				

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/788,966

Applicant(s)

PINCHUK ET AL.

Examiner

Brian E. Pellegrino

Art Unit

3738

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 21 February 2006.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 31-39 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 31-39 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 24 March 2006 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☐ Other: _____

DETAILED ACTION

Claim Rejections - 35 USC § 102

The text of those sections of Title 35, U.S. Code not included in this action can be found in a prior Office action.

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 31,32,34 are rejected under 35 U.S.C. 102(b) as anticipated by or, in the alternative, under 35 U.S.C. 103(a) as obvious over Herweck et al. (5197976). Fig. 4 shows a tubular component that has been creased or longitudinally indented to form a double lumen.

However, in the alternative Herweck does not explicitly disclose creasing the tubular component. Herweck et al. does disclose (col. 8, lines 1-3) that the tubular component is perforated, which is understood that this is done to separate portions of the tube into multiple components or lumens. It is well known in the art to use molding apparatus to press into tubular components to crease the pliable structure when forming or extruding PTFE, such that a single extruded tube can be used to form a multiple lumen structure. It would have been obvious to one of ordinary skill in the art to utilize creasing structure to form a multiple lumen structure from the tubular component of Herweck such that it forms symmetrical lumens.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 33,35 are rejected under 35 U.S.C. 103(a) as being unpatentable over Herweck et al. '976 in view of Harbuck (4501263). Herweck et al. is explained supra. However, Herweck fails to disclose a triple lumen or quadruple lumen tubular component. Harbuck discloses a triple lumen component for use in a hepatic area of repair needing a three lumen portion graft component, Fig. 1. It would have been obvious to one of ordinary skill in the art to utilize the teaching of Harbuck in forming a triple lumen structure with the prosthetic device of Herweck et al. such that it could be used in a hepatic repair surgical procedure. Harbuck also teaches a plurality of branches could be used, col. 3, lines 16-19, col. 5, lines 13-15 such that it would have been an obvious to one of ordinary skill in the art to use four lumens as suggested by Harbuck with the prosthesis of Herweck because if there are four branches off of a main vessel that are diseased or collapsed vessels not functioning it would have been an obvious expedient to utilize a quadruple lumen portion to repair the vessels.

Double Patenting

The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. A nonstatutory obviousness-type double patenting rejection is appropriate where the conflicting claims are not identical, but at least one examined application claim is not patentably distinct from the reference claim(s) because the examined application claim is either anticipated by, or would have been obvious over, the reference claim(s). See, e.g., *In re Berg*, 140 F.3d 1428, 46 USPQ2d 1226 (Fed. Cir. 1998); *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) or 1.321(d) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent either is shown to be commonly owned with this application, or claims an invention made as a result of activities undertaken within the scope of a joint research agreement.

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

Claims 36-39 are rejected on the ground of nonstatutory obviousness-type double patenting as being unpatentable over claims 1,2,5 of U.S. Patent No. 6554858. Although the conflicting claims are not identical, they are not patentably distinct from each other because this application's claims are so close in content that they both cover the same thing, despite a slight difference in wording than the claims of 6554858. Regarding claim 38, self-expanding stents are well known in the art.

Response to Arguments

Applicant's arguments filed 2/21/06 have been fully considered but they are not persuasive. Regarding the arguments about the double patenting rejection over US '858, Applicant argues that claim 1 is directed to a "circular cross-section along its entire length.." and since claim 36 of the current application only requires "a portion of the trunk to be non-indented" it can be said to read on it. Additionally, claim 1 of US patent '858 recites the "sleeve having diametrically opposite portions which are joined to one another in a manner to form the axial channels" that the Examiner considers to be analogous to the claimed "indents which join portions of the sleeve to form the channels" in current application claim 36. Thus the rejection is maintained.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be

Art Unit: 3738

calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Brian E. Pellegrino whose telephone number is 571-272-4756. The examiner can normally be reached on Monday-Thursday from 7am to 4:30pm. The examiner can also be reached on alternate Fridays.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Corrine McDermott, can be reached on 571-272-4754. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

TC 3700, AU 3738

A handwritten signature in cursive script that reads "Brian E. Pellegrino".

BRIAN E. PELLEGRINO
PRIMARY EXAMINER